

*Facility Name*  
*date*  
**Uniformity of Screen Speed**

Screen Type:

Processor:

Film Type:

18 x 24 Emulsion #:

24 x 30 Emulsion#:

Imaging mode:

Density Control Setting:

Focal spot size:

kVp setting:

Cassette  
I.D. #

Image I.D.  
#

Recorded  
mAs

Measured  
Optical  
Density

18 x 24 cm

_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____

Maximum Optical Density: 0.00  
Minimum Optical Density: 0.00  
Optical Density Range: 0.00

24 x 30 cm

_____	_____
_____	_____
_____	_____
_____	_____
_____	_____
_____	_____

Maximum Optical Density: 0.00  
Minimum Optical Density: 0.00  
Optical Density Range: 0.00

**Action Limit:** Uniformity of screen speed of all cassettes shall be tested and the difference between the minimum and maximum optical densities shall not exceed 0.30. If the values exceed 0.30 corrective action is needed.